

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME PARVOVIRUS B19 IgG - CHORUS

Code	81150				
Lot No.	926		2024-04		2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D0220	Parvovirus B19 IgG	926	2024-04	6X6
B0220	CALIBRATORE Parvovirus B19 IgG	926	2024-04	1
C0220	CONTROL SERUM Parvovirus B19 IgG	926	2024-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	302	≥ 100
Control Serum (Index)	3,9	2,9-5,4

CALIBRATOR

0	1	1	9	2	6	6	0	0	0	0	0	2	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

0	1	1	9	2	6	6	2	9	5	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *HB*

DATE *06-05-2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.