

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Pag 1/1
Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS VITAMIN EXTRACTION

Code 86500

Lot No. 002

 2025-09

 2023-04

REAGENTS

Label	Description	Lot.	Exp	No
B4580	Releasing Agent	002	2025-12	1
C4580	Stabilizing Agent	002	2025-09	1
D4580	Neutralizing Agent	002	2025-12	1

CONTROLS

	Found results	Expected results
pH Releasing Agent	13,2	> 13,0
pH Stabilizing Agent	<2,0	< 2,0
pH Neutralizing Agent	4,21	4,0-4,5

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 05.05.2023