

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-G - CHORUS

Code 86100

Lot No. 926



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	926	2024-04	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	926	2024-11	1
C8660	CONTROL SERUM Deamidated Gliadin-G	926	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	410	> 100
Control Serum (AU/mL)	116,0	46,1-188,9

CALIBRATOR

2 5 5 9 2 6 6 1 1 9 2 6 3 8 4 4 0 0

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*




CONTROL SERUM

2 5 5 9 2 6 6 1 8 3 4 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *[Signature]*

DATE 10.05.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.