

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39  
Tel +39-0577-319554 Fax +39-0577-319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

### QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-W5 WORMWOOD

|         |          |  |         |  |         |
|---------|----------|--|---------|--|---------|
| Code    | 88000-W5 |  |         |  |         |
| Lot No. | 926      |  | 2024-04 |  | 2023-05 |

### REAGENTS

| Label | Description                     | Lot | Exp     | No  |
|-------|---------------------------------|-----|---------|-----|
| D4991 | 88000-W5 WORMWOOD               | 926 | 2024-04 | 2X6 |
| B4990 | CALIBRATORE 88000-W5 WORMWOOD   | 926 | 2024-06 | 1   |
| C4990 | CONTROL SERUM 88000-W5 WORMWOOD | 926 | 2024-06 | 1   |

### CONTROLS

|  | Found results | Expected result |
|--|---------------|-----------------|
|--|---------------|-----------------|

|                      |      |           |
|----------------------|------|-----------|
| Calibrator (mOD)     | 279  | ≥ 100     |
| Control Serum (KU/L) | 17,5 | 10,1-44,7 |

### CALIBRATOR


|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 2 | 6 | 9 | 2 | 6 | 6 | 5 | 2 | 0 | 6 | 2 | 1 | 8 | 7 | 4 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature .....



### CONTROL SERUM


|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 2 | 6 | 9 | 2 | 6 | 6 | 3 | 9 | 7 | 2 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature .....



The test has been carried out in accordance with the package insert

TESTED BY: Francesco Bisconti

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 09.05.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.