

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SM - CHORUS

Code	86016		
Lot No.	919	2024-06	2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3161	SM	919	2024-06	6X6
B3161	CALIBRATORE SM	919	2024-06	1
C3161	CONTROL SERUM SM	919	2024-06	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	669	> 100
Control Serum (AU/mL)	51,3	26,2-60,0

CALIBRATOR

1	9	2	9	1	9	7	4	1	4	2	4	5	5	3	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	9	2	9	1	9	7	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 11-05-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.