

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG -CHORUS

Code	81010				
Lot No.	920		2024-03		2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2161	Cytomegalovirus IgG	920	2024-03	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	920	2024-03	1
C2161	CONTROL SERUM Cytomegalovirus IgG	920	2024-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	309	≥ 100
Control Serum (IU/mL)	11,3	7,7-20,5
Negative Control (IU/mL)	<0,50	<0,5

CALIBRATOR

0	0	1	9	2	0	6	5	2	4	3	5	2	3	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	0	1	9	2	0	6	3	0	5	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 12/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.