

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgM - CHORUS

Code 81041

Lot No. 930



2024-04



2023-04

### REAGENTS

| Label | Description                  | Lot. | Exp.    | NO  |
|-------|------------------------------|------|---------|-----|
| D2281 | TOXOPLASMA IgM               | 930  | 2024-04 | 6X6 |
| B2281 | CALIBRATORE TOXOPLASMA IgM   | 930  | 2025-04 | 1   |
| C2281 | CONTROL SERUM TOXOPLASMA IgM | 930  | 2025-04 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                          |     |            |
|--------------------------|-----|------------|
| Calibrator (mOD)         | 400 | $\geq 100$ |
| Control Serum (INDEX)    | 2,2 | 1,5-2,9    |
| Negative Control (INDEX) | 0,1 | $< 0,7$    |


### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 3 | 1 | 9 | 3 | 0 | 6 | 0 | 0 | 0 | 0 | 0 | 4 | 6 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature *RB*




### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 3 | 1 | 9 | 3 | 0 | 6 | 1 | 5 | 2 | 9 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature *RB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 16.05.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.