

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr EBNA IgG -CHORUS

Code 81057

Lot No. 934



2024-05



2023-05

### REAGENTS

Label	Description	Lot.	Exp.	NO
D2361	Epstein-Barr EBNA IgG	934	2024-05	6X6
B2362	CALIBRATORE Epstein-Barr EBNA IgG	934	2024-05	1
C2361	CONTROL SERUM Epstein-Barr EBNA IgG	934	2024-05	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	430	$\geq 100$
Control Serum (INDEX)	2,3	1,8-3,3
Negative Control (INDEX)	0,0	$<0,7$

### CALIBRATOR

0	0	6	9	3	4	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



### CONTROL SERUM

0	0	6	9	3	4	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

*[Handwritten signature]*

DATE 29/05/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.