

QUALITY CONTROL CERTIFICATE

KIT NAME a-TG - CHORUS

Code 86072

Lot No. 918



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3571	a-TG	918	2024-03	6X6
B3571	CALIBRATORE a-TG	918	2024-03	1
C3571	CONTROL SERUM a-TG	918	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	368	> 100
Control Serum (IU/mL)	487,2	281,8-666,7


CALIBRATOR

3	0	0	9	1	8	6	1	2	7	4	5	3	4	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*




CONTROL SERUM

3	0	0	9	1	8	6	2	1	3	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....*[Signature]*

DATE 29-05-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.