

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgA - CHORUS

Code 81252

Lot No. 939 2024-05 2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D8610	Chlamydomphila pneumoniae IgA	939	2024-05	6X6
B8610	CALIBRATORE Chlamydomphila p. IgA	939	2024-05	1
C8610	CONTROL SERUM Chlamydomphila p. IgA	939	2024-05	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	344	≥ 100
Control Serum (AU/mL)	47,0	29,2-70,0
Negative Control (AU/mL)	<10,0	≤ 10,0

CALIBRATOR

1	9	9	9	3	9	6	8	1	3	2	6	3	8	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

1	9	9	9	3	9	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE 31.05.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.