

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgM CHORUS

Code	81076/12				
Lot No.	926		2024-04		2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2461	MUMPS IgM	926	2024-04	2x6
B2461	CALIBRATORE MUMPS IgM	926	2024-04	1
C2461	CONTROL SERUM MUMPS IgM	926	2024-04	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	442	≥ 100
Control Serum (Index)	2,0	1,6-3,0
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR

1	3	2	9	2	6	6	0	0	0	0	0	4	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	3	2	9	2	6	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 31.05.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.