

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME **tTg-A - CHORUS**

Code	86058		
Lot No.	936	2024-05	2023-05

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	936	2024-05	6X6
B3431	CALIBRATORE tTg-A	936	2024-05	1
C3431	CONTROL SERUM tTg-A	936	2024-05	1


### CONTROLS

	Found results	Expected results
Calibrator (mOD)	762	$\geq 100$
Control Serum (AU/mL)	32,0	22,5-54,5

### CALIBRATOR

3	3	9	9	3	6	6	4	2	0	3	1	4	6	3	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



### CONTROL SERUM

3	3	9	9	3	6	6	1	5	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 13.06.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.