

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA URINARY ANTIGEN - CHORUS

Code 81301

Lot No. 927 2024-01 2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D9530	LEGIONELLA URINARY ANTIGEN	927	2024-01	2X6
B9530	CALIBRATOR LEGIONELLA URINARY ANTIGEN	927	2024-01	1
C9530	POSITIVE CONTROL LEGIONELLA URINARY ANTIGEN	927	2024-01	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	240	≥ 100
Positive Control (Index)	3,9	2,6-4,9

CALIBRATOR

2	7	0	9	2	7	5	0	0	0	0	0	2	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

2	7	0	9	2	7	5	2	6	4	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Blus*

DATE 15/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.