

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-G - CHORUS

Code 86056

Lot No. 928



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	928	2024-04	6X6
B3411	CALIBRATORE Gliadin-G	928	2024-04	1
C3411	CONTROL SERUM Gliadin-G	928	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	569	≥ 100
CONTROL SERUM (AU/ml)	74,4	36,0-130,8

CALIBRATOR


5	0	2	9	2	8	6	3	2	0	2	7	4	5	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

5	0	2	9	2	8	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 12/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.