

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code 86600

Lot No. 943



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	943	2024-06	2 X 6
B9500	CALIBRATORE CALPROTECTIN	943	2024-08	1
C9500	CONTROL CALPROTECTIN	943	2024-08	1
G9502	SAMPLE DILUENT	943	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	450	≥ 100
Control (mg/kg)	133,1	75,0-207,1

CALIBRATOR

9	0	0	9	4	3	6	2	1	6	2	0	5	4	2	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RB



CONTROL SERUM

9	0	0	9	4	3	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 15/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.