Pag 1/1

DIESSE DIAGNOSTICA SENESE SpA

53035 Monteriggioni (SI) – Italia – Via delle Rose 10 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione 20144 Milano (Italy) Via A. Solari 19 sc. 6 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

PRODUCT NAME:

VACU-TEC

Code	10602					
Lot No	3109	\square	2024-02	~	2023-04	
Reference	ce IO-09/88					

ON LINE QUALITY CONTROLS (model CDP not enclosed)			QC Passed	
•	sodium citrate solution concentration		~	
	sodium citrate volume		ℯ	
•	vacuum level (*)		₽	
•	correct placement of the cap		₩	
	no extra body inside the product		F	
(*)	Only for VACU-CODE and VACU-TEC tubes		₽	
PA	CKAGING QUALITY CONTROL (model CDP not enclosed)		QC Passed	
	control of the number of the product for each pack		ত	
QU	ALITY CONTROLS ON FINISHED PRODUCT (model CDP	not enclosed)	QC Passed	
	pack integrity		7	
	data printed on the product label	9 *	₽	
			QC Passed	N/A
Irra	diation treatment Certificate n° 4048/2023 date 09/06/20	23	₽ 2	Γ
			QC Passed	N/A
	mma rays treatment process in compliance with the spe he supply Agreement	cifications	ᅜ	٣
Dod	cument number:	ODV 137989		

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.

DATE 14/06/2023