

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel: +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI IgG -CHORUS**

Code **81060**

Lot No. **943**



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	943	2024-06	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	943	2024-06	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	943	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	394	≥ 100
Control Serum (UA/mL)	87,8	59,4-144,4
Negative Control (UA/mL)	<5	<5

CALIBRATOR


0	0	8	9	4	3	6	2	2	0	3	6	3	0	5	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM

0	0	8	9	4	3	6	2	3	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Alma*

DATE *16.06.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.