



DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

## QUALITY CONTROL CERTIFICATE

KIT NAME 88000-T3 COMMON SILVER BIRCH

Code	88000-T3		
Lot No.	930	2024-04	2023-06

### REAGENTS

Label	Description	Lot.	Exp.	NO
D4971	88000-T3 COMMON SILVER BIRCH	930	2024-04	2X6
B4970	CALIBRATORE 88000-T3 COMMON SILVER BIRCH	930	2024-06	1
C4970	CONTROL SERUM 88000-T3 COMMON SILVER BIRCH	930	2024-06	1

### CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	279	> 100
Control Serum (KU/L)	17,5	10,1-44,7

### CALIBRATOR

6	0	5	9	3	0	6	5	2	0	6	2	1	8	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



### CONTROL SERUM

6	0	5	9	3	0	6	3	9	7	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 23-06-2023

#### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.