

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MPO - CHORUS

Code 86068

Lot No. 936



2024-05



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3531	MPO	936	2024-05	6X6
B3531	CALIBRATORE MPO	936	2024-05	1
C3531	CONTROL SERUM MPO	936	2024-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	526	> 100
Control Serum (AU/mL)	50,2	30,0-74,3

CALIBRATOR

2	5	3	9	3	6	6	1	1	8	2	8	3	9	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Operator Signature



CONTROL SERUM

2	5	3	9	3	6	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Signature*

DATE 29.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.