

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BORRELIA IgG - CHORUS

Code	81046		
Lot No.	952	2024-07	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2300	BORRELIA IgG	952	2024-07	6X6
B2300	CALIBRATORE BORRELIA IgG	952	2024-09	1
C2300	CONTROL BORRELIA IgG	952	2024-09	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	400	≥ 100
Control (Index)	3,1	2,0-3,7

CALIBRATOR

9	0	4	9	5	2	6	0	0	0	0	0	3	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	0	4	9	5	2	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 05/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.