

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgM

Code 81059

Lot No. 943



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2382	Epstein-Barr Early Antigen IgM	943	2024-06	6x6
B2382	CALIBRATORE Epstein-Barr Early Antigen IgM	943	2024-06	1
C2382	CONTROL SERUM Epstein-Barr Early Antigen IgM	943	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (Index)	3,3	2,3-4,3
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR


0	2	9	9	4	3	6	0	0	0	0	0	3	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM


0	2	9	9	4	3	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alvoro*

DATE *11/07/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.