

QUALITY CONTROL CERTIFICATE

KIT NAME POLIOVIRUS IgG - CHORUS

Code 81264/12

Lot No. 940



2024-05



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D9630	Poliovirus IgG	940	2024-05	2X6
B9630	CALIBRATORE Poliovirus IgG	940	2024-06	1
C9630	CONTROL SERUM Poliovirus IgG	940	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	546	> 100
Control (Index)	1,5	1,3-2,3

CALIBRATOR

9	0	1	9	4	0	6	0	0	0	0	0	5	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

9	0	1	9	4	0	6	1	3	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 26.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.