

QUALITY CONTROL CERTIFICATE

KIT NAME INFLUENZA B IgG -CHORUS

Code 81193

Lot No. 952



2024-07



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D7531	INFLUENZA B IgG	952	2024-07	6X6
B7531	CALIBRATORE INFLUENZA B IgG	952	2024-07	1
C7531	CONTROL SERUM INFLUENZA B IgG	952	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	4,35	> 100
Control Serum (Ndx)	2,6	1,7-3,1
Negative Control (Ndx)	0,0	≤ 0,7


CALIBRATOR

0	7	2	9	5	2	6	0	0	0	0	0	4	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

0	7	2	9	5	2	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 20.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.