

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME POLIOVIRUS IgG - CHORUS

Code 81264

Lot No. 940



2024-05



2023-06

### REAGENTS

Label	Description	Lot.	Exp.	NO
D9630	Poliovirus IgG	940	2024-05	6X6
B9630	CALIBRATORE Poliovirus IgG	940	2024-06	1
C9630	CONTROL SERUM Poliovirus IgG	940	2024-06	1

### CONTROLS

	Found results	Expected results
Calibrator (mOD)	546	> 100
Control (Index)	1,5	1,3-2,3

### CALIBRATOR


9	0	1	9	4	0	6	0	0	0	0	0	5	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Ben*



### CONTROL SERUM


9	0	1	9	4	0	6	1	3	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Ben*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Ben*

DATE 24-07-2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.