

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgM CHORUS

Code 81079

Lot No. 952



2024-10



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2481	VARICELLA IgM	952	2024-10	6x6
B2481	CALIBRATORE VARICELLA IgM	952	2024-10	1
C2481	CONTROL SERUM VARICELLA IgM	952	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	620	> 100
Control Serum (Index)	2,2	1,7-3,1
Negative Control (Index)	0,1	< 0,7


CALIBRATOR

1	3	4	9	5	2	7	0	0	0	0	0	4	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*




CONTROL SERUM

1	3	4	9	5	2	7	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....*[Signature]*

DATE 24.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.