

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code	86094
Lot No.	945
	2024-06
	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
DB671	ANTI-CCP	945	2024-06	6x6
BB671	CALIBRATORE ANTI-CCP	945	2024-06	1
CB671	CONTROL SERUM ANTI-CCP	945	2024-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	510	≥ 100
Control Serum (AU/mL)	73,7	52,6-100,0

CALIBRATOR

2	5	6	9	4	5	6	3	2	0	3	9	3	4	5	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *B*



CONTROL SERUM

2	5	6	9	4	5	6	2	1	3	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *B*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *B*

DATE 26/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.