

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MPO - CHORUS

Code	86068		
Lot No.	952	2024-07	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3531	MPO	952	2024-07	6X6
B3531	CALIBRATORE MPO	952	2024-07	1
C3531	CONTROL SERUM MPO	952	2024-07	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	479	≥ 100
Control Serum (AU/mL)	45,7	27,9-65,7

CALIBRATOR

2	5	3	9	5	2	6	3	1	7	2	7	4	6	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2	5	3	9	5	2	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 28.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.