

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgG

Code	81075		
Lot No.	936	2024-08	2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2451	MUMPS IgG	936	2024-08	6X6
B2452	CALIBRATORE MUMPS IgG	936	2024-08	1
C2451	CONTROL SERUM MUMPS IgG	936	2024-08	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control Serum (Index)	3.0	2.2-4.0
Negative Control (Index)	0.0	< 0,7

CALIBRATOR

0	1	0	9	3	6	7	0	0	0	0	0	3	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature MB



CONTROL SERUM

0	1	0	9	3	6	7	2	2	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature MB



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... MB

DATE 28/07/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.