Code

Lot No

Pag 1/1

DIESSE DIAGNOSTICA SENESE SpA

53035 Monteriggioni (SI) – Italia – Via delle Rose 10 Tel +39-0577-587111 Fax +39-0577-318690



2024-05

Sede Legale e Amministrazione 20144 Milano (Italy) Via A. Solari 19 sc. 6 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

2023-07

PRODUCT NAME:

10200/W

3206

**VACU-TEC** 

Refe	erence IO-09/88		
ON LINE QUALITY CONTROLS (model CDP not enclosed)		QC Passed	
-	sodium citrate solution concentration	F	
-	sodium citrate volume	₽.	
•	vacuum level (*)	ਹ	
-	correct placement of the cap	₽	
=	no extra body inside the product	ᄝ	
(*) C	only for VACU-CODE and VACU-TEC tubes	ᅜ	
PAC	KAGING QUALITY CONTROL (model CDP not enclosed)	QC Passed	
•	control of the number of the product for each pack	V	
QUA	ALITY CONTROLS ON FINISHED PRODUCT (model CDP not enclosed)	QC Passed	
M.	pack integrity	च	
	data printed on the product label	⊽	
		QC Passed	N/A
Irrac	liation treatment Certificate n° date	Г	₽
		QC Passed	N/A
	nma rays treatment process in compliance with the specifications ne supply Agreement	Г	₽
Document number: ODV N.A.			

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

**APPROVED** 

RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 28/07/2023