

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgM-CHORUS

Code	81011
Lot No.	928
	2024-07
	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2171	Cytomegalovirus IgM	928	2024-07	6X6
B2171	CALIBRATORE Cytomegalovirus IgM	928	2024-07	1
C2171	CONTROL SERUM Cytomegalovirus IgM	928	2024-07	1
G2171	CONFIRMATORY TEST	928	2024-07	5X1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	537	≥ 100
Control Serum (INDEX)	2,2	1,5-2,8
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0	7	9	9	2	8	7	0	0	0	0	0	0	4	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	7	9	9	2	8	7	1	5	2	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 09.08.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.