

QUALITY CONTROL CERTIFICATE

KIT NAME **HERPES SIMPLEX 1+2 IgG -CHORUS**

Code **81020**

Lot No. **937**



2024-08



2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2211	Herpes simplex 1+2 IgG	937	2024-08	6X6
B2212	CALIBRATORE Herpes simplex 1+2 IgG	937	2024-08	1
C2211	CONTROL SERUM Herpes simplex 1+2 IgG	937	2024-08	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	345	≥ 100
Control Serum (INDEX)	2,7	2,1-4,0
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0	0	4	9	3	7	7	0	0	0	0	0	3	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

0	0	4	9	3	7	7	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE **28-08-2023**

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.