

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code	81270/12		
Lot No.	944	2024-06	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	944	2024-06	2X6
B9660	CALIBRATORE Tetanus IgG	944	2024-07	1
C9660	CONTROL SERUM Tetanus IgG	944	2024-07	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	522	≥ 100
Control (mIU/mL)	638,1	428,6-1000,0

CALIBRATOR

9	0	2	9	4	4	6	2	1	7	2	8	4	7	5	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	0	2	9	4	4	6	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 23/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.