

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

Code 86900

Lot No. 929



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	929	2024-04	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	929	2025-04	1
C9703	CONTROL SERUM 25OH Vitamin D Total	929	2025-04	4

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	454	≥ 100
CONTROL SERUM (ng/ml)	37,3	33,1-39,3

CALIBRATOR


8	3	5	9	2	9	6	1	3	6	4	1	6	5	6	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

8	3	5	9	2	9	6	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 27/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.