

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgM - CHORUS

Code 81041

Lot No. 934

2024-05

2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D2281	TOXOPLASMA IgM	934	2024-05	6X6
B2281	CALIBRATORE TOXOPLASMA IgM	934	2024-08	1
C2281	CONTROL SERUM TOXOPLASMA IgM	934	2024-08	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	396	≥ 100
Control Serum (INDEX)	1,9	1,6-2,9
Negative Control (INDEX)	0,1	$< 0,7$


CALIBRATOR

1	3	1	9	3	4	6	0	0	0	0	0	4	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

1	3	1	9	3	4	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 31/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.