

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA IgA - CHORUS

Code 81033

Lot No. 952



2024-07



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D7551	MYCOPLASMA IgA	952	2024-07	6X6
B7551	CALIBRATORE MYCOPLASMA IgA	952	2024-07	1
C7551	CONTROL SERUM MYCOPLASMA IgA	952	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	452	≥ 100
Control Serum (AU/mL)	41,5	26,7-48,2
Negative Control (AU/mL)	<10,0	<10

CALIBRATOR

0	6	8	9	5	2	6	7	1	2	2	1	4	6	7	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	6	8	9	5	2	6	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 04/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.