

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI IgA-CHORUS**

Code **81062**

Lot No. **961**

 **2024-08**

 **2023-08**

REAGENTS

Label	Description	Lot.	Exp.	NO
D2401	Helicobacter Pylori IgA	961	2024-08	6X6
B2401	CALIBRATORE Helicobacter Pylori IgA	961	2024-08	1
C2401	CONTROL SERUM Helicobacter Pylori IgA	961	2024-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (INDEX)	3,4	1,9-3,6
Negative Control (INDEX)	0,0	$< 0,7$


CALIBRATOR

0 2 7 9 6 1 6 0 0 0 0 3 7 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature 



CONTROL SERUM

0 2 7 9 6 1 6 1 9 3 6 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE **01-08-2023**

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.