

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel, +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgM CHORUS

Code 81073

Lot No. 943



2024-09



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2442	MEASLES IgM	943	2024-09	6x6
B2442	CALIBRATORE MEASLES IgM	943	2024-09	2
C2442	CONTROL SERUM MEASLES IgM	943	2024-09	2

CONTROLS

	Found results	Expected results
Calibrator (mOD)	460	≥ 100
Control Serum (Index)	2,6	1,9-3,5
Negative Control (index)	0,0	$< 0,7$


CALIBRATOR

1	3	3	9	4	3	7	0	0	0	0	0	3	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *AP*




CONTROL SERUM

1	3	3	9	4	3	7	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *AP*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *AP*

DATE 04.09.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.