

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code	81035
Lot No.	947
	2024-09
	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2261	MYCOPLASMA PNEUMONIAE IgM	947	2024-09	6x6
B2261	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	947	2024-10	1
C2261	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	947	2024-10	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	430	≥ 100
Control Serum (INDEX)	2,4	1,6-3,0
Negative Control (INDEX)	0,1	$\leq 0,7$

CALIBRATOR

0	6	7	9	4	7	7	0	0	0	0	0	4	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature HB



CONTROL SERUM

0	6	7	9	4	7	7	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Deine

DATE 05/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive