

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-A - CHORUS

Code 86054

Lot No. 945



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3391	Gliadin-A	945	2024-06	6X6
B3392	CALIBRATORE Gliadin-A	945	2024-06	1
C3392	CONTROL SERUM Gliadin-A	945	2024-06	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	300	≥ 100
CONTROL SERUM (AU/ml)	63,7	42,0-130,0

CALIBRATOR

5	0	1	9	4	5	6	5	3	2	4	7	2	6	6	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *RB*



CONTROL SERUM

5	0	1	9	4	5	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *AB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *AB*

DATE 05/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.