

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CLOSTRIDIUM DIFFICILE GDH - CHORUS

Code	81168		
Lot No.	958	2024-07	2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D1880	CLOSTRIDIUM DIFFICILE GDH	958	2024-07	2X6
B1880	CALIBRATORE C.DIFFICILE GDH	958	2024-07	1
C1880	CONTROL SERUM C.DIFFICILE GDH	958	2024-07	1
G1880	C.DIFFICILE Stool Diluent/Negative Control	D0144500	2024-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	370	≥ 100
Control + (Index)	2,7	1,8-3,4
Stool Diluent/Negative Control (Index)	0,1	< 0.5

CALIBRATOR

9	3	5	9	5	8	6	0	0	0	0	0	3	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature HB



CONTROL SERUM

9	3	5	9	5	8	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents HB

DATE 07/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!