

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgG - CHORUS

Code 81250

Lot No. 953



2024-07



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D8590	Chlamydomphila pneumoniae IgG	953	2024-07	6X6
B8590	CALIBRATORE Chlamydomphila p. IgG	953	2024-07	1
C8590	CONTROL SERUM Chlamydomphila p. IgG	953	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	≥ 100
Control Serum (AU/mL)	38,7	24,3-57,1
Negative Control (AU/mL)	<10,0	$\leq 10,0$

CALIBRATOR


2	2	2	9	5	3	6	7	1	5	2	9	4	0	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	2	2	9	5	3	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 13-09-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.