

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghetti
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-G - CHORUS

Code 86056

Lot No. 955



2024-07



2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	955	2024-07	6X6
B3411	CALIBRATORE Gliadin-G	955	2024-08	1
C3411	CONTROL SERUM Gliadin-G	955	2024-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	730	≥ 100
CONTROL SERUM (AU/ml)	93,0	35,0-161,5

CALIBRATOR

5	0	2	9	5	5	6	2	1	7	2	3	4	9	3	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *SB*



CONTROL SERUM

5	0	2	9	5	5	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Blue*

DATE *13.09.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious reagents should be handled with maximum attention.