

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code 81021

Lot No. 955

2024-07

2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2221	Herpes simplex 1+2 IgM	955	2024-07	6x6
B2221	CALIBRATORE Herpes simplex 1+2 IgM	955	2024-07	1
C2221	CONTROL SERUM Herpes simplex 1+2 IgM	955	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	480	≥ 100
Control Serum (INDEX)	2,9	2,0-3,7
Negative Control (INDEX)	0,1	$< 0,7$

CALIBRATOR

1	5	6	9	5	5	6	0	0	0	0	0	3	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Blum*



CONTROL SERUM


1	5	6	9	5	5	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Blum*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Blum*

DATE 31-08-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.