

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-A - CHORUS

Code 86098

Lot No. 962



2024-08



2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D8650	Deamidated Gliadin-A	962	2024-08	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	962	2024-08	1
C8650	CONTROL SERUM Deamidated Gliadin-A	962	2024-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	412	≥ 100
Control Serum (AU/mL)	254,9	79,2-400,0

CALIBRATOR

2	5	4	9	6	2	6	2	2	4	3	6	2	3	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

83



CONTROL SERUM

2	5	4	9	6	2	6	3	1	5	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

83



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

M. Lisa

DATE 22/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.