

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr EBNA IgG -CHORUS

Code 81057

Lot No. 970



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2361	Epstein-Barr EBNA IgG	970	2024-09	6X6
B2362	CALIBRATORE Epstein-Barr EBNA IgG	970	2024-09	1
C2361	CONTROL SERUM Epstein-Barr EBNA IgG	970	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	340	≥ 100
Control Serum (INDEX)	2,5	1,9-3,4
Negative Control (INDEX)	0,0	$<0,7$

CALIBRATOR

0 0 6 9 7 0 6 0 0 0 0 0 3 8 0 0 0 0


Label checked with barcode reader

YES



Operator Signature

HB



CONTROL SERUM

0 0 6 9 7 0 6 1 9 3 4 0 0 0 0 0 0 1


Label checked with barcode reader

YES



Operator Signature

HB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

HB

DATE 28/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.