

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CLOSTRIDIUM DIFFICILE A/B TOXINS - CHORUS

Code 81170

Lot No. 970

2024-06

2023-09

REAGENTS

Label	Description	Lot	Exp.	NO
D1890	CLOSTRIDIUM DIFFICILE A/B TOXINS	970	2024-06	2x6
B1890	CALIBRATORE C.DIFFICILE A/B TOXINS	970	2024-07	1
C1890	CONTROL SERUM C.DIFFICILE A/B TOXINS	970	2024-07	1
G1890	C.DIFFICILE Stool Diluent/Negative Control	C0312701	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	410	≥ 100
Control + (Index)	3,5	2,4-4,5
Stool Diluent/Negative Control (Index)	0,2	$< 0,5$


CALIBRATOR

9	3	6	9	7	0	5	0	0	0	0	0	2	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature



CONTROL SERUM

9	3	6	9	7	0	5	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

29/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.