

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	954	2024-07	2023-07

## REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	954	2024-07	6X6
B3451	CALIBRATORE tTg-G	954	2024-07	1
C3451	CONTROL SERUM tTg-G	954	2024-07	1

## CONTROLS

	Found results	Expected results
Calibrator (mOD)	464	≥ 100
Control Serum (AU/mL)	36,9	30,0-55,7

## CALIBRATOR

3	4	0	9	5	4	6	1	1	3	2	7	5	3	3	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature HB



## CONTROL SERUM

3	4	0	9	5	4	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HB



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... Alcer

DATE 06/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive