

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgG - CHORUS

Code 81040

Lot No. 970



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2271	TOXOPLASMA IgG	970	2024-09	6X6
B2272	CALIBRATORE TOXOPLASMA IgG	970	2024-09	1
C2271	CONTROL SERUM TOXOPLASMA IgG	970	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	> 100
Control Serum (IU/mL)	28,5	20,0-40,0
Negative Control (IU/mL)	<5	<5

CALIBRATOR

0	8	2	9	7	0	6	3	2	5	4	5	3	5	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Orlen*



CONTROL SERUM

0	8	2	9	7	0	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Orlen*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Orlen*

DATE 19-10-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.