

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 2 IgG RECOMBINANT - CHORUS

Code 81029

Lot No. 972



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D8690	Herpes simplex 2 IgG Recombinant	972	2024-09	6X6
B8690	CALIBRATORE Herpes simplex 2 IgG Recombinant	972	2024-09	1
C8690	CONTROL SERUM Herpes simplex 2 IgG Recombinant	972	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	610	≥ 100
Control Serum (INDEX)	1,6	1,2-2,3
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

2 6 0 9 7 2 6 0 0 0 0 0 5 8 0 0 0 0

Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

2 6 0 9 7 2 6 1 2 2 3 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature *HO*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *HB*

DATE 16/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.