

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Chorus Chlamydia trachomatis IgA

Code 81262

Lot No. 970



2024-09



2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D9650	Chorus Chlamydia trachomatis IgA	970	2024-09	6X6
B9650	CALIBRATOR Chlamydia trachomatis IgA	970	2024-09	1
C9650	POSITIVE CONTROL Chlamydia trachomatis IgA	970	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	410	≥ 100
Positive Control (Index)	2,6	1,8-3,3

CALIBRATOR

2	3	2	9	7	0	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



CONTROL SERUM

2	3	2	9	7	0	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Amg*

DATE *17/10/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.